

DICOMscope - Bug #527

Print server not quitting on shutdown

2013-06-27 15:22 - Michael Onken

Status:	Closed	Start date:	2013-06-27
Priority:	Normal	Due date:	
Assignee:		% Done:	100%
Category:		Estimated time:	0:00 hour
Target version:	3.6.2		
Description			
Print server is not quitting on shutdown.			

History

#1 - 2019-03-29 15:39 - Marco Eichelberg

- Status changed from New to Resolved
- % Done changed from 0 to 100

This was caused by expired X.509 certificates and only affected print server processes operating over TLS.
Fixed in DICOMscope 3.6.4 (at least for the next 10 years).

#2 - 2020-05-25 13:49 - Michael Onken

- Status changed from Resolved to Closed