

## DCMTK - Issues

| #    | Tracker     | Status   | Priority | Subject                                                                                                                               | Assignee       | Updated          |
|------|-------------|----------|----------|---------------------------------------------------------------------------------------------------------------------------------------|----------------|------------------|
| 1133 | Conformance | New      | Normal   | The dcmdata parser should check private creator elements for forbidden characters                                                     |                | 2024-08-16 14:16 |
| 1017 | Conformance | New      | Normal   | Check handling of enhanced multiframe images by compression codecs                                                                    |                | 2022-03-04 17:39 |
| 1000 | Conformance | New      | Normal   | Add support for CP-2083: Add uncompressed encapsulated Transfer Syntax for encoding very large pixel data sets–                       | Jörg Riesmeier | 2024-06-26 18:14 |
| 985  | Conformance | New      | Normal   | Add support for CP-1984: Add Slide Orientation and Total Pixel Matrix information to Segmentation and Parametric Maps for TILED_FULL– | Michael Onken  | 2021-04-30 17:15 |
| 980  | Conformance | New      | Normal   | Add support for CP-2041: Add Table Value Type for DICOM SR                                                                            | Jörg Riesmeier | 2024-10-30 16:55 |
| 977  | Conformance | Resolved | Normal   | Add support for Supplement 214: Enhanced X-Ray Radiation Dose SR (including Cone-beam CT)                                             | Jörg Riesmeier | 2021-11-12 12:18 |
| 955  | Conformance | New      | Normal   | Add support for CP-2041: Add Table Value Type for DICOM SR                                                                            | Jörg Riesmeier | 2020-11-18 09:12 |
| 931  | Conformance | New      | Normal   | Remove redundant mapping of DIMSE Status Codes to human readable description                                                          |                | 2020-04-17 10:48 |
| 869  | Conformance | New      | Normal   | Check whether "undefined length" encoding is supported for all VRs it is specified                                                    |                | 2020-05-25 13:29 |
| 866  | Conformance | New      | Normal   | Add support for CP-1818: Large compressed images may have more frames than fit in the Basic Offset Table–                             |                | 2019-01-25 16:15 |
| 776  | Conformance | New      | Normal   | Ensure multiple value matching is implemented correctly.                                                                              |                | 2018-01-31 11:10 |
| 726  | Conformance | New      | Normal   | CP-1649: Clarify use of Other Patient IDs Sequence with Issuer and related Attributes as Query keys–                                  |                | 2020-06-02 22:07 |
| 695  | Conformance | Assigned | Normal   | Check whether Non-Patient Object Storage SOP Classes should be treated differently                                                    | Jörg Riesmeier | 2020-05-25 13:29 |
| 685  | Conformance | New      | Low      | CP-1539 extends the "Enhanced Encoding Mode" (Code Sequence Macro)                                                                    | Jörg Riesmeier | 2016-05-31 15:15 |
| 645  | Conformance | New      | Normal   | Check whether CP-1411 (Add attributes describing safe or unsafe private elements) requires any changes                                |                | 2015-07-02 16:38 |
| 642  | Conformance | New      | Normal   | Add support for Supplement 186: Extensible SR Storage SOP Class                                                                       | Jörg Riesmeier | 2016-10-27 13:32 |
| 624  | Conformance | Assigned | Normal   | CP-1031 introduces Long and URN Code Values                                                                                           | Jörg Riesmeier | 2016-10-27 13:35 |
| 141  | Conformance | New      | Normal   | check whether PixelPaddingValue should be used for modality/VOI transform                                                             |                | 2015-07-02 16:35 |